

Robotics and Automation in Veterinary Surgery

Actuator – Concept: device that converts electrical, hydraulic, or pneumatic energy into mechanical motion. Related terms: servo motor, pneumatic cylinder, linear actuator. Explanation: Actuators drive robotic arms, end-effectors, and positioning platforms in veterinary surgery. Example: a pneumatic actuator controls the opening of a laparoscopic port. Practical application: precise manipulation of surgical instruments in minimally invasive procedures. Challenge: ensuring smooth, repeatable motion without generating excessive heat that could affect surrounding tissue.

Artificial Intelligence (AI) – Concept: computational techniques that enable machines to mimic human reasoning and learning. Related terms: machine learning, deep learning, neural networks. Explanation: AI algorithms process intra-operative imaging, sensor data, and surgeon inputs to guide robotic assistance. Example: AI-based vision systems identify anatomical landmarks in a canine thorax. Practical application: real-time decision support for instrument selection. Challenge: obtaining sufficient labeled veterinary data to train robust models.

Autonomous Robot – Concept: a robot capable of performing tasks without continuous human control. Related terms: semi-autonomous system, supervisory control, decision-making algorithm. Explanation: In veterinary surgery, autonomous robots may execute pre-programmed suturing sequences. Example: an autonomous robot places a series of interrupted sutures on a feline intestinal anastomosis. Practical application: reducing surgeon fatigue during lengthy procedures. Challenge: guaranteeing safety when unexpected tissue variability occurs.

Biomechanical Modeling – Concept: computational representation of animal musculoskeletal structures. Related terms: finite element analysis, kinematic simulation, tissue elasticity. Explanation: Models predict forces on instruments and tissues during robotic manipulation. Example: a 3-D finite-element model of a horse's stifle joint informs optimal force thresholds for a robotic drill. Practical application: customizing robot parameters for different species. Challenge: acquiring accurate material properties for diverse veterinary tissues.

Calibration – Concept: process of aligning robot sensors and actuators to known standards. Related terms: zeroing, pose verification, tool tracking. Explanation: Calibration ensures that commanded movements correspond to actual instrument tip locations. Example: laser-based calibration aligns a robotic arm's coordinate system with the surgical table. Practical application: maintaining sub-millimeter accuracy across multiple surgeries. Challenge: repeatable calibration in a busy clinic where environmental conditions change.

Computer-Aided Design (CAD) – Concept: software tools for creating precise 3-D models. Related terms: CAM, STL file, parametric modeling. Explanation: CAD models of surgical instruments and patient anatomy guide robot path planning. Example: a CAD model of a canine femur is imported into the robot's planning software. Practical application: rapid prototyping of custom end-effectors. Challenge: integrating

veterinary-specific anatomical libraries into mainstream CAD platforms.

Computer-Vision – Concept: techniques that enable machines to interpret visual data. Related terms: image segmentation, object detection, stereo vision. Explanation: In veterinary surgery, computer-vision algorithms track instrument tips and tissue deformation. Example: a convolutional neural network distinguishes between healthy and inflamed gastric tissue in a rabbit. Practical application: feedback for adaptive force control. Challenge: coping with variable lighting, fur, and blood that obscure visual cues.

Control Loop – Concept: feedback system that continuously adjusts robot motion based on sensor input. Related terms: PID controller, closed-loop control, sensor fusion. Explanation: The loop maintains desired trajectory and force during instrument insertion. Example: a PID controller modulates a robotic scalpel's speed to prevent tearing of delicate skin. Practical application: achieving smooth cutting while preserving tissue integrity. Challenge: tuning parameters for different animal sizes and tissue types.

Cyber-Physical System (CPS) – Concept: integration of computation, networking, and physical processes. Related terms: Internet of Things, real-time monitoring, edge computing. Explanation: A veterinary surgical robot is a CPS that synchronizes sensor data, control algorithms, and actuation. Example: real-time telemetry from force sensors streams to a cloud-based analytics platform. Practical application: remote monitoring of robot health and performance. Challenge: ensuring data security and low latency in a clinical environment.

Deep Learning – Concept: subset of machine learning using multi-layer neural networks. Related terms: convolutional neural network, backpropagation, training dataset. Explanation: Deep learning models classify tissue types, predict instrument trajectories, and detect complications. Example: a CNN predicts the likelihood of postoperative infection from intra-operative images. Practical application: automated alerts to the surgeon. Challenge: avoiding over-fitting due to limited veterinary case numbers.

Denavit-Hartenberg (DH) Parameters – Concept: standardized method for describing robot geometry. Related terms: kinematic chain, joint axis, transformation matrix. Explanation: DH parameters define the relative positions of robot links and joints for path planning. Example: a 6-DOF veterinary robot's link offsets are expressed using DH conventions. Practical application: simplifying inverse kinematics calculations. Challenge: adapting generic DH models to custom-built veterinary end-effectors.

Dexterity – Concept: ability of a robot to manipulate objects with precision and flexibility. Related terms: degrees of freedom, articulation, end-effector design. Explanation: High dexterity enables delicate dissection of soft tissues in small animal surgery. Example: a wristed instrument replicates a surgeon's hand motions during a feline ophthalmic procedure. Practical application: expanding the range of procedures that can be robot-assisted. Challenge: balancing dexterity with instrument size to avoid excessive bulk.

End-Effector – Concept: device attached to the robot's arm that interacts with the surgical site. Related terms: gripper, scalpel, ultrasonic probe. Explanation: End-effectors can be interchangeable tools such as suturing devices or biopsy needles. Example: a micro-gripper designed for avian beak surgery. Practical application: rapid tool changes between steps of a multi-phase operation. Challenge: sterilization and material compatibility with diverse veterinary species.

Force Feedback (Haptics) – Concept: system that conveys tactile information from the robot to the surgeon. Related terms: tactile sensor, vibrotactile actuator, teleoperation. Explanation: Haptic feedback alerts the surgeon to tissue resistance, preventing over-compression. Example: a haptic joystick vibrates when a robotic trocar contacts the abdominal wall of a dog. Practical application: enhancing surgeon confidence in remote or minimally invasive settings. Challenge: achieving high bandwidth feedback without latency.

Finite Element Analysis (FEA) – Concept: computational method for predicting how structures respond to forces. Related terms: mesh generation, stress-strain analysis, material model. Explanation: FEA evaluates bone stress during robotic drilling in large animal orthopedics. Example: an FEA model predicts cortical thinning risk when a robotic burr engages a horse's metacarpal. Practical application: informing safe cutting parameters. Challenge: generating accurate meshes from heterogeneous veterinary imaging data.

Four-Bar Linkage – Concept: simple planar mechanism composed of four rigid bars connected by pivots. Related terms: kinematic synthesis, transmission angle, mechanical advantage. Explanation: Used in compact robotic manipulators for small animal surgeries. Example: a miniature four-bar mechanism drives a micro-scissor for feline skin closure. Practical application: providing precise angular motion in a limited space. Challenge: scaling the mechanism while preserving stiffness.

Gyroscope – Concept: sensor that measures angular velocity. Related terms: inertial measurement unit, accelerometer, sensor fusion. Explanation: Gyroscopes help maintain orientation of a robot's end-effector during complex maneuvers. Example: a MEMS gyroscope detects subtle wrist rotations of a robotic arm during a laparoscopic ovariectomy. Practical application: compensating for table movement. Challenge: drift errors that accumulate over long procedures.

Hybrid Robot – Concept: system combining multiple actuation modalities (e.g., electric and pneumatic). Related terms: dual-mode actuation, compliance control, modular architecture. Explanation: Hybrid robots can switch between high-speed positioning and compliant force control. Example: an electric-driven arm for rapid approach, then pneumatic compliance for delicate tissue retraction. Practical application: optimizing speed and safety across surgical phases. Challenge: synchronizing control algorithms across actuation domains.

Image-Guided Surgery (IGS) – Concept: use of real-time imaging to direct surgical tools. Related terms: CT navigation, intra-operative ultrasound, fluoroscopy. Explanation: Robots integrate IGS to align instrument trajectories with patient anatomy. Example: a robotic drill follows a pre-planned path based on a CT scan of a canine skull. Practical application: improving accuracy of tumor resections. Challenge: managing radiation exposure and registration errors.

Inertial Measurement Unit (IMU) – Concept: sensor suite combining accelerometers, gyroscopes, and sometimes magnetometers. Related terms: sensor fusion, orientation tracking, drift compensation. Explanation: IMUs provide pose data for robotic arms, especially when optical tracking is obstructed by fur. Example: an IMU attached to a robotic instrument monitors its orientation during a cat's spinal surgery. Practical application: maintaining spatial awareness in cluttered operative fields. Challenge: calibrating against external references to reduce cumulative error.

Joint Space – Concept: representation of robot configuration using joint variables (angles, displacements). Related terms: task space, inverse kinematics, configuration manifold. Explanation: Planning in joint space simplifies collision avoidance for multi-DOF veterinary robots. Example: a joint-space trajectory avoids the robot's own arm colliding with the animal's limb. Practical application: generating smooth motion profiles. Challenge: handling singularities that arise in certain configurations.

Kinematics – Concept: study of motion without regard to forces. Related terms: forward kinematics, inverse kinematics, Jacobian matrix. Explanation: Kinematic models predict end-effector position from joint angles and vice versa. Example: solving inverse kinematics to position a robotic needle tip at a specific depth in a rabbit's liver. Practical application: precise targeting of lesions. Challenge: computational load for real-time adjustments in dynamic surgical environments.

Laser Ablation – Concept: removal of tissue using focused laser energy. Related terms: photothermal effect, fiber optic delivery, laser safety. Explanation: Robotic platforms can precisely steer laser fibers for selective tissue removal. Example: a robot-guided CO₂ laser excises a pigmented skin lesion on a horse's flank. Practical application: minimizing blood loss and collateral damage. Challenge: controlling heat diffusion in highly vascular tissues.

Learning Curve – Concept: rate at which proficiency improves with practice. Related terms: skill acquisition, competency assessment, proficiency plateau. Explanation: Adoption of robotic surgery in veterinary practice requires training to overcome initial performance deficits. Example: a surgeon's operative time halves after ten robot-assisted spays. Practical application: designing curricula that accelerate skill mastery. Challenge: variability in prior experience among veterinary surgeons.

Linear Actuator – Concept: device that produces straight-line motion. Related terms: lead screw, ball screw, rack-and-pinion. Explanation: Linear actuators control insertion depth of tools such as biopsy needles. Example: a motor-driven leadscrew advances a trocar into a canine abdomen with millimeter precision. Practical application: repeatable depth control across cases. Challenge: maintaining force feedback while moving rapidly.

Machine Vision Calibration – Concept: process of aligning camera coordinates with robot coordinates. Related terms: extrinsic calibration, intrinsic parameters, checkerboard pattern. Explanation: Accurate calibration ensures that visual detections correspond to real-world locations. Example: a checkerboard mounted on the surgical table calibrates a robot's stereo cameras before a feline laparoscopy. Practical application: reliable instrument tracking. Challenge: re-calibrating after any change in camera pose or lighting.

Manipulator – Concept: robot arm that positions and orients tools. Related terms: kinematic chain, degrees of freedom, payload capacity. Explanation: Veterinary manipulators may be designed for specific species size ranges. Example: a 7-DOF manipulator accommodates both small-animal (rabbit) and large-animal (dog) procedures. Practical application: flexible platform for a mixed practice. Challenge: balancing reach, stiffness, and compactness.

Medical Imaging Modality – Concept: technique for visualizing internal anatomy. Related terms: CT, MRI,

ultrasound, fluoroscopy. Explanation: Imaging informs robot path planning and intra-operative navigation. Example: an ultrasound-guided robot places a catheter in a horse's jugular vein. Practical application: reducing blind insertions. Challenge: integrating heterogeneous data streams in real time.

Micro-Robot – Concept: miniature robot designed for extremely small operative fields. Related terms: nanorobotics, MEMS, micro-actuator. Explanation: Enables procedures on tiny patients such as rodents or exotic birds. Example: a micro-robotic arm performs microsutures on a finch's wing tendon. Practical application: advancing research models and rare-species care. Challenge: fabricating reliable components at sub-millimeter scales.

Motion Planning – Concept: algorithmic determination of a collision-free path for the robot. Related terms: RRT, A* algorithm, trajectory optimization. Explanation: Motion planners compute safe routes around animal anatomy and surgical tools. Example: an RRT planner generates a path for a robotic drill to avoid the femoral artery in a dog. Practical application: automated path generation reduces surgeon workload. Challenge: accounting for tissue deformation that occurs during surgery.

Neural Network – Concept: interconnected layers of artificial neurons that learn patterns. Related terms: deep learning, backpropagation, activation function. Explanation: Neural networks predict optimal force thresholds based on prior surgeries. Example: a network suggests suction pressure for a cat's thoracic cavity based on intra-operative vitals. Practical application: adaptive control. Challenge: avoiding black-box decisions that are difficult to interpret clinically.

Operative Field – Concept: physical space where surgery is performed. Related terms: sterile barrier, access port, workspace constraints. Explanation: Robotic systems must fit within the limited operative field of small animals. Example: a compact robot arm fits inside a 10 cm diameter drape for a rabbit abdomen. Practical application: preserving sterility while providing robotic reach. Challenge: designing tools that do not obstruct the surgeon's view.

Optical Tracker – Concept: camera system that detects markers to locate objects in space. Related terms: infrared markers, motion capture, registration. Explanation: Tracks instrument tips and animal anatomy to maintain alignment with robot commands. Example: an optical tracker follows reflective markers on a canine limb during orthopedic robot-assisted plating. Practical application: dynamic registration during bone manipulation. Challenge: marker occlusion by fur or surgical staff.

Patient-Specific Modeling – Concept: creation of individualized anatomical models from imaging. Related terms: segmentation, 3-D reconstruction, personalized planning. Explanation: Tailors robot trajectories to each animal's unique geometry. Example: a 3-D model of a horse's femur guides custom drill trajectories. Practical application: improving accuracy and reducing intra-operative adjustments. Challenge: time-consuming segmentation for complex anatomies.

Pedicle Screw Insertion Robot – Concept: specialized robot for placing screws in vertebral pedicles. Related terms: spinal navigation, drill guide, torque control. Explanation: In large-animal spinal surgery, the robot aligns the drill to a pre-planned trajectory. Example: a robot places six pedicle screws in a camel's lumbar spine. Practical application: enhancing screw placement accuracy. Challenge: accounting for variable bone

density across species.

Perception System – Concept: suite of sensors that interpret the environment. Related terms: vision, force, tactile, audio. Explanation: Provides the robot with data on tissue properties, instrument location, and surgeon commands. Example: a combined camera-force system detects when a robotic gripper contacts soft tissue. Practical application: enabling adaptive control strategies. Challenge: fusing heterogeneous sensor data in real time.

PID Controller – Concept: control algorithm using proportional, integral, and derivative terms. Related terms: feedback loop, tuning, set-point. Explanation: Maintains desired position or force during robotic manipulation. Example: a PID controller regulates the speed of a robotic scalpel to keep cutting force constant. Practical application: smooth, predictable motion. Challenge: selecting gains that work across a range of animal tissue stiffnesses.

Planar Robot – Concept: robot whose motion is confined to a two-dimensional plane. Related terms: Cartesian robot, gantry system, X-Y table. Explanation: Useful for surface procedures such as dermatological laser treatment. Example: a planar robot moves a laser head across a dog's back for uniform hair removal. Practical application: high-throughput treatments. Challenge: limited to flat work surfaces, requiring repositioning for curved anatomy.

Portable Surgical Robot – Concept: compact, lightweight robot designed for use outside a fixed operating suite. Related terms: battery-powered, modular, field-deployable. Explanation: Enables robotic assistance in field clinics or mobile veterinary units. Example: a portable robot assists in spaying stray cats in a rural outreach program. Practical application: extending advanced care to underserved areas. Challenge: maintaining performance with limited power and environmental control.

Precision Medicine – Concept: tailoring treatment to individual biological characteristics. Related terms: genomics, personalized dosing, targeted therapy. Explanation: Robotic surgery can deliver precise interventions aligned with precision medicine goals. Example: a robot places a drug-eluting implant in a dog's tumor based on molecular profiling. Practical application: maximizing therapeutic effect while minimizing side effects. Challenge: integrating genomic data with real-time surgical planning.

Proprioceptive Sensor – Concept: sensor that measures internal robot states such as joint angles or motor currents. Related terms: encoders, strain gauges, torque sensors. Explanation: Provides feedback on robot posture without external measurement. Example: joint encoders report the exact angle of a robotic wrist during a feline eye surgery. Practical application: enabling closed-loop control without line-of-sight. Challenge: sensor drift over long procedures.

Programmable Logic Controller (PLC) – Concept: industrial computer used for automation and control. Related terms: ladder logic, I/O modules, real-time execution. Explanation: PLCs manage peripheral devices such as suction pumps and cautery units in a veterinary robot suite. Example: a PLC synchronizes robot motion with electrosurgical activation. Practical application: reliable coordination of multiple subsystems. Challenge: ensuring interoperability with medical-grade software standards.

Qualitative Kinematic Analysis – Concept: assessment of robot motion based on descriptive criteria rather

than numerical metrics. Related terms: smoothness, fluency, ergonomic rating. Explanation: Evaluates surgeon comfort and workflow during robot-assisted procedures. Example: surgeons rate the fluidity of a robot's arm movements during a series of canine castrations. Practical application: informing design improvements. Challenge: subjectivity and variability among users.

Real-Time Operating System (RTOS) – Concept: software platform that guarantees timing constraints for critical tasks. Related terms: deterministic scheduling, latency, priority inversion. Explanation: RTOS ensures that sensor data, control loops, and safety checks execute within strict deadlines. Example: an RTOS schedules force-feedback updates at 1 kHz during a rabbit thoracotomy. Practical application: maintaining safety and responsiveness. Challenge: integrating RTOS with higher-level AI modules that may have variable compute times.

Remote Center of Motion (RCM) – Concept: constraint that forces a tool to pivot about a fixed point, mimicking a trocar entry. Related terms: cannula, entry point, kinematic constraint. Explanation: RCM mechanisms allow robotic instruments to enter the body without violating the skin incision. Example: a robotic arm with an RCM passes through a 5 mm port for a feline laparoscopic ovariectomy. Practical application: preserving incision integrity. Challenge: achieving precise RCM behavior while maintaining flexibility.

Robot-Assisted Endoscopy – Concept: use of robotic platforms to navigate endoscopic scopes. Related terms: flexible endoscope, navigation algorithm, image stabilization. Explanation: Improves maneuverability in narrow cavities such as the equine sinus. Example: a robot steers a flexible endoscope through a horse's nasal passage to reach a sinus cyst. Practical application: reducing manual fatigue and enhancing view stability. Challenge: dealing with variable lumen compliance and secretions.

Robot Kinematics Calibration – Concept: refinement of the mathematical model linking joint parameters to end-effector pose. Related terms: parameter identification, error mapping, calibration phantom. Explanation: Ensures that commanded positions match actual positions within tolerances. Example: a calibration routine uses a grid of known points to adjust DH parameters for a canine orthopedic robot. Practical application: maintaining accuracy across multiple surgeries. Challenge: time-consuming procedures that must be performed regularly.

Safety Interlock – Concept: hardware or software mechanism that prevents unsafe robot operation. Related terms: emergency stop, watchdog timer, fault detection. Explanation: Stops motion if unexpected forces exceed thresholds or if a sensor fails. Example: an interlock disables the robot when the force sensor detects a sudden spike indicating possible tissue rupture. Practical application: protecting both patient and staff. Challenge: designing interlocks that are fail-safe yet do not cause unnecessary interruptions.

Scalability – Concept: ability of a robotic system to adapt to different sizes and species. Related terms: modular design, adjustable reach, payload variation. Explanation: A scalable robot can be configured for a rabbit or a horse by changing tool lengths and control parameters. Example: interchangeable linear actuators allow the same platform to handle both small-animal and large-animal surgeries. Practical application: cost-effective investment for mixed practices. Challenge: ensuring performance does not degrade at extreme size extremes.

Sensor Fusion – Concept: integration of data from multiple sensors to produce a more accurate estimate. Related terms: Kalman filter, complementary filter, data fusion algorithm. Explanation: Combines vision, force, and IMU data to determine tool tip location. Example: a Kalman filter merges camera coordinates with force sensor readings during a canine laparoscopic suturing task. Practical application: robust position tracking despite occlusions. Challenge: balancing computational load with real-time requirements.

Servo Motor – Concept: motor that provides precise control of angular position, velocity, and torque. Related terms: closed-loop control, encoder feedback, PWM drive. Explanation: Drives the joints of most veterinary surgical robots. Example: a high-torque servo rotates the wrist of a robotic instrument for fine suture placement in a cat's intestine. Practical application: accurate, repeatable joint movements. Challenge: heat dissipation in compact enclosures.

Six-Degree-of-Freedom (6-DOF) Robot – Concept: robot capable of moving in three translational and three rotational axes. Related terms: spatial manipulation, Cartesian robot, articulated arm. Explanation: Provides full spatial freedom required for complex veterinary procedures. Example: a 6-DOF robot positions a drill tip to a precise angle on a horse's tibia. Practical application: versatility across many surgical specialties. Challenge: increased programming complexity and collision risk.

Simulation Environment – Concept: virtual platform for testing robot behavior before physical deployment. Related terms: physics engine, virtual reality, digital twin. Explanation: Allows surgeons to rehearse procedures on patient-specific models. Example: a surgeon practices a canine spinal fixation in a simulated environment that mirrors the real robot's dynamics. Practical application: reducing intra-operative errors. Challenge: achieving high fidelity representations of soft tissue deformation.

Soft Robotics – Concept: robots built from compliant materials that can adapt to irregular shapes. Related terms: pneumatic artificial muscle, elastomeric actuator, bio-inspired design. Explanation: Soft grippers can gently handle delicate tissues such as avian skin. Example: a silicone-based soft manipulator wraps around a chicken's wing without causing bruising. Practical application: minimizing trauma in fragile patients. Challenge: controlling precise force output while maintaining compliance.

Surgical Navigation System – Concept: technology that tracks instruments relative to patient anatomy. Related terms: registration, tracking, augmented reality. Explanation: Provides the robot with spatial context for tool placement. Example: a navigation system registers a CT scan of a dog's skull, allowing the robot to drill precisely for a cranial tumor excision. Practical application: enhancing accuracy of minimally invasive techniques. Challenge: maintaining registration accuracy as the animal moves or tissue deforms.

Telemetry – Concept: remote transmission of data from sensors to a monitoring station. Related terms: wireless communication, data logging, real-time monitoring. Explanation: Sends robot status, force readings, and video streams to a central console. Example: force telemetry alerts the surgeon when a robotic instrument exceeds safe pressure during a feline liver biopsy. Practical application: continuous oversight of robot performance. Challenge: ensuring reliable connectivity in metal-rich OR environments.

Tool Changeover – Concept: process of swapping one end-effector for another during a procedure. Related terms: quick-release interface, modular mount, sterile dock. Explanation: Enables a single robot to perform

multiple tasks such as cutting, suturing, and cauterizing. Example: an automated tool changer replaces a scalpel with a bipolar cautery tip mid-operation on a dog's intestine. Practical application: reducing downtime between steps. Challenge: guaranteeing sterility and alignment after each change.

Trajectory Optimization – Concept: mathematical refinement of a robot's path to minimize criteria like time, energy, or tissue stress. Related terms: cost function, gradient descent, dynamic programming. Explanation: Produces smoother, safer motions for delicate surgeries. Example: optimizing a drilling trajectory to avoid high-stress regions in a horse's femur. Practical application: enhancing patient outcomes and reducing operative time. Challenge: incorporating real-time tissue deformation into the optimization loop.

Ultrasound-Guided Robotic Intervention – Concept: use of ultrasound imaging to direct robot actions. Related terms: sonic tracking, needle placement, acoustic coupling. Explanation: Allows percutaneous procedures without ionizing radiation. Example: a robot inserts a biopsy needle into a canine liver under continuous ultrasound guidance. Practical application: safe sampling of deep organs. Challenge: maintaining image quality despite probe movement and acoustic shadowing.

Virtual Fixtures – Concept: software-generated constraints that guide the surgeon's movements. Related terms: haptic guidance, forbidden region, assistive overlay. Explanation: Virtual fixtures can restrict a robotic instrument to a safe corridor. Example: a virtual cylinder prevents a robot's scalpel from deviating outside the intended incision line during a cat's skin graft. Practical application: reducing inadvertent tissue damage. Challenge: designing intuitive fixtures that do not impede necessary freedom.

Vision-Based Servoing – Concept: control strategy that uses visual feedback to regulate robot motion. Related terms: image-based control, visual servo loop, feature tracking. Explanation: Adjusts robot pose to keep a target feature centered in the camera view. Example: a robot tracks the tip of a suture needle as it advances through a dog's abdominal wall. Practical application: compensating for patient movement. Challenge: latency and robustness to variable lighting.

Wireless Power Transfer – Concept: delivering energy to robot components without physical connectors. Related terms: inductive coupling, resonant charging, battery-less operation. Explanation: Enables untethered robotic tools in sterile fields. Example: a wireless power pad supplies energy to a small robotic gripper used in a rabbit eye surgery. Practical application: eliminating cables that could interfere with the surgical field. Challenge: ensuring sufficient power while avoiding electromagnetic interference with monitoring equipment.

Zero-Force Positioning – Concept: robot positioning mode where no external force is applied to the environment. Related terms: gravity compensation, compliant mode, passive compliance. Explanation: Allows the surgeon to manually move the robot's arm without resistance, useful for teaching or repositioning. Example: a zero-force mode lets a veterinarian position a robotic arm over a horse's limb before activating active control. Practical application: intuitive hand-over-hand teaching. Challenge: maintaining stability when transitioning back to active control.

Zoom Lens Imaging – Concept: optical system that provides variable magnification for intra-operative cameras. Related terms: telecentric lens, focal length, resolution scaling. Explanation: Allows surgeons to

view both the overall operative field and fine details without swapping cameras. Example: a zoom lens captures a wide view of a canine thorax while allowing close-up inspection of a suture line. Practical application: flexible visualization. Challenge: maintaining focus and depth of field across magnification changes.

Z-Axis Control – Concept: precise management of vertical movement in a Cartesian robot. Related terms: linear actuator, vertical positioning, depth control. Explanation: Critical for inserting needles or drills to exact depths. Example: a Z-axis controller positions a biopsy needle 12 mm beneath the skin of a rabbit's flank. Practical application: repeatable depth accuracy across procedures. Challenge: compensating for tissue compression that alters effective depth.